

NT-proBNP

(EN) ENZYME IMMUNOASSAY FOR THE QUANTITATIVE DETERMINATION OF NT-proBNP
IN HUMAN SERUM AND EDTA PLASMA
CAT. NO. SK-1204 12 X 8 TESTS

(DE) ENZYME IMMUNOASSAY FÜR DIE QUANTITATIVE BESTIMMUNG VON NT-proBNP
IN HUMANEN SERUM UND EDTA PLASMA PROBEN
KAT. NR. SK-1204 12 X 8 TESTE

FOR RESEARCH USE ONLY
NOT FOR USE IN DIAGNOSTIC PROCEDURES

rev.no. 190111 (replacing 181206)

This kit was developed and manufactured by:
Biomedica Medizinprodukte GmbH, A-1210 Wien, Divischgasse 4
Tel. +43/1/291 07 45, Fax +43/1/291 07 6389, E-mail info@bmgrp.com



CONTENT / INHALT / SOMMAIRE / CONTENUTO

ENGLISH	3
DEUTSCH	7

Additional information on our products is available on our website.
Zusätzliche Information zu unseren Produkten ist auf unserer Homepage erhältlich.

www.bmgrp.com

1) INTRODUCTION

BNP is mainly expressed by ventricular myocardium in response to volume overload and increased filling pressure. BNP has a cleavable signal sequence. Mature BNP consists of 108 amino acids (proBNP or BNP-108), and undergoes cleavage resulting in physiologically active BNP-32 and additional C-terminal fragments (cf. http://www.uniprot.org/uniprot/P16860#PRO_000001532), along with a physiologically inactive N-terminal peptide comprising amino acids 1-76, which is further degraded proteolytically.

BNP has a key role in cardiovascular homeostasis with biological actions including natriuresis, diuresis, vasorelaxation, and inhibition of renin and aldosterone secretion. A high concentration of BNP in the bloodstream is indicative of heart failure.

Areas of Interest

- Cardiac impairment, acute myocardial infarction, (left ventricular dysfunction)
- Renal failure
- Obesity and diabetes
- Various forms of secondary hypertension
- Therapy monitoring of heart failure patients

2) CONTENTS OF THE KIT

CONT	KIT COMPONENTS	QUANTITY
PLATE	polyclonal sheep anti NT-proBNP antibody coated microtiter strips in stripholder packed in alu bag with desiccant	12 x 8 tests
WASHBUF	Wash buffer, 20x concentrated, clear cap	1 x 50 ml
STD	Standards, synthetic human NT-proBNP (0, 10, 40, 160, 640 pmol/l), lyophilised, white caps	5 vials
CTRL	Controls, synthetic human NT-proBNP, lyophilised, yellow cap, exact concentrations after reconstitution see labels	2 vials
CONJ	Conjugate, (sheep anti human NT-proBNP-HRPO), red dye, brown cap, ready to use	1 x 22 ml
SUB	Substrate (TMB solution), blue cap, ready to use	1 x 22 ml
STOP	Stop solution, sulphuric acid, white cap, ready to use	1 x 7 ml

3) ADDITIONAL MATERIAL ADDED TO THE KIT

- 1 self-adhesive plastic film
- QC protocol
- Protocol sheet
- Instruction manual for use

4) MATERIAL AND EQUIPMENT REQUIRED BUT NOT SUPPLIED

- Precision pipettes calibrated to deliver 50-500 µl and disposable tips
- ELISA reader for absorbance at 450 nm (630 nm reference optional)
- Graph paper or software for calculation of results
- Plate washer is recommended for washing, alternatively: multichannel pipette or dispenser
- Refrigerator with 4°C (2-8°C)
- Distilled or deionised water

5) REAGENTS AND SAMPLE PREPARATION

All reagents as supplied in the kit are stable at 4°C (2-8°C) until the expiry date stated on the label of each reagent.

Sample preparation:

NT-proBNP is stable in whole blood for several hours at room temperature (18-26°C). Nevertheless we recommend separating EDTA plasma or serum by centrifugation as soon as possible, e.g. 20 min at 2,000 x g, preferably at 4°C (2-8°C). EDTA plasma or serum can be stored at 4°C (2-8°C) up to two days. For long term storage, aliquot the acquired samples and store them at -25°C or lower. Samples can be subjected to 3 freeze-thaw cycles. Lipemic or hemolyzed samples may give erroneous results. Samples should be mixed well before assaying. We recommend duplicates for all values. Samples with values above STD5 (640 pmol/l) can be diluted with STD1 or NT-proBNP negative human serum.


For further information on sample stability and assay performance characteristics please visit our website www.bmgrp.com (s. Validation Data) or contact our customer service by e-mail info@bmgrp.com or by phone +43/ 1/ 29107-45.

Reconstitution/Handling:

- **STD (Standard):** Pipette 500 µl of distilled or deionised water into each vial. Leave at room temperature (18-26°C) for 10 min. Swirl gently. The standard concentration is printed on the label. Reconstituted standard is stable at -25°C until expiry date. Avoid freeze-thaw cycles.
- **CTRL (Control):** Pipette 500 µl of distilled or deionised water to each vial. Leave at room temperature (18-26°C) for 10 min. Swirl gently. The final concentration is stated on the label. Reconstituted control is stable at -25°C until expiry date stated on label. Avoid freeze-thaw cycles.
- **WASHBUF (Wash buffer):** Dilute the concentrate 1:20 (e.g. 50 ml WASHBUF + 950 ml distilled water). Crystals in the buffer concentrate will dissolve at room temperature (18-26°C). The undiluted WASHBUF is stable at 4°C (2-8°C) until expiry date stated on label. The diluted WASHBUF is stable at 4°C (2-8°C) up to one month. Use only diluted WASHBUF (Wash buffer) for the assay.

6) PRINCIPLE OF THE ASSAY

This kit is a sandwich enzyme immunoassay for the determination of NT-proBNP in human serum or EDTA plasma. In a first step, sample and conjugate (sheep anti human NT-proBNP-HRPO) are pipetted into the wells of the microtiter strips, which are pre-coated with polyclonal sheep anti NT-proBNP antibody. NT-proBNP present in the sample binds to the pre-coated antibody in the well and forms a sandwich with the conjugate (detection antibody). In the washing step all non-specific unbound material is removed. In a second step, the substrate (TMB Tetramethylbenzidine) is pipetted into the wells. The enzyme catalysed colour change of the substrate is directly proportional to the amount of NT-proBNP present in the sample. This colour change is detectable with a standard microtiter plate ELISA reader.



7) ASSAY PROTOCOL

All reagents and samples must be at room temperature (18-26°C) before use in the assay.

Mark position for STD/SAMPLE/CTRL (Standard/Sample/Control) on the protocol sheet.


Take microtiter strips out of the alu bag. Store unused strips with desiccant at 4°C (2-8°C) in the alu bag. Strips are stable until expiry date stated on the label.

1. Add 50 µl STD/SAMPLE/CTRL (Standards/Sample/Control) in duplicate into respective well.
2. Add 200 µl CONJ (Conjugate) into each well, swirl gently.
3. **Cover tightly and incubate 3 hours at room temperature (18-26°C).**
4. Aspirate and wash wells 5x with 300 µl diluted WASHBUF (Wash buffer), remove remaining WASHBUF by strongly tapping plate against paper towel after the last wash.
5. Add 200 µl SUB (Substrate) into each well, swirl gently.
6. **Incubate for 30 min at room temperature (18-26°C) in the dark.**
7. Add 50 µl STOP (Stop solution) into each well, swirl gently.
8. Measure absorbance immediately at 450 nm with reference 630 nm, if available.

8) CALCULATION OF RESULTS

Read the optical density (OD) of all wells on a plate reader using 450 nm wavelength (correction wavelength 630 nm). Construct the standard curve from the OD values of the STD. Use commercially available software or graph paper. Obtain sample concentration from this standard curve. The assay was evaluated with 4PL algorithm. Different curve fitting methods need to be evaluated by the user. Samples with values above STD5 (640 pmol/l) can be diluted with STD1 or NT-proBNP negative human serum and re-assayed. Dilution factors must be considered when calculating the final sample concentrations.

Example typical STD-curve:



The quality control (QC) protocol supplied with the kit shows the results of the final release QC for each kit at production date. Data for OD obtained by customers may differ due to various influences and/or due to the normal decrease of signal intensity during shelf life. However, this does not affect validity of results as long as an OD of 1.50 or more is obtained for STD5 and the values of the CTRLs are in range (target ranges see labels).

9) ASSAY CHARACTERISTICS

Method:	Sandwich ELISA, HRP/TMB, 12x8-well strips		
Antibodies/Standard:	Capture antibody: polyclonal sheep anti human NT-proBNP antibody, specifically binding to amino acids 31-57 of proBNP. Detection antibody: polyclonal sheep anti human NT-proBNP antibody, conjugated to peroxidase, specifically binding to amino acids 8-29 of proBNP. Standard material: synthetic human NT-proBNP (1-76).		
Sample type:	Human serum, EDTA plasma		
Standard range:	0 to 640 pmol/l (0/10/40/160/640 pmol/l)		
Conversion factor pmol/l to pg/ml	1 pmol/l = 8.475 pg/ml refers to NT-proBNP (1-76) that is detected by the ELISA		
Sample volume:	50 µl / well		
Incubation time:	3 hours / 30 min		
Sensitivity:	LOD (0 pmol/l + 3SD): 3.0 pmol/l; LLOQ: 3.3 pmol/l		
Specificity:	This assay recognizes endogenous (natural) and recombinant human NT-proBNP (1-76).		
Precision:	Intra-assay (n=3) ≤ 4%, Inter-assay (n=8) ≤ 7%		
Spike/Recovery (average R spiked with rec. NT-proBNP)	Spike 80 pmol/l		Spike 320 pmol/l
	Serum (n=4) 99%		108%
	EDTA Plasma (n=4) 94%		93%
Dilution linearity (average R of expected NT-proBNP after a 1+1 dilution)	Serum (rec. NT-proBNP): 117%		Serum (endog. NT-proBNP): 82%
	EDTA plasma (rec. NT-proBNP): 84%		EDTA plasma (endog. NT-proBNP): 84%
Values of apparently healthy individuals:	Median (serum, n=70): 5.8 pmol/l Median (EDTA plasma, n=28): 5.6 pmol/l Each laboratory should establish its own reference data.		

For further information on assay characteristics please visit our website www.bmgrp.com (see Validation Data) or contact our customer service by e-mail info@bmgrp.com or by phone +43/ 1/ 29107-45.

10) PRECISION

Experiment:

Intra-assay: 2 samples of known concentrations were tested 3 times in 1 assay by 1 operator.

Inter-assay: 2 samples of known concentrations were tested 8 times in 2 assays by different operators.

Intra-assay (n=3)	Sample 1	Sample 2	Inter-assay (n=8)	Sample 1	Sample 2
Mean (pmol/l)	60.2	35.2	Mean (pmol/l)	52.1	108.1
SD (pmol/l)	2.0	0.9	SD (pmol/l)	1.7	7.9
CV (%)	4	3	CV (%)	3	7

11) TECHNICAL HINTS

- Do not mix or substitute reagents with those from other lots or sources.
- Do not mix stoppers and caps from different reagents or various lots.
- Do not use reagents beyond expiration date. Protect reagents from direct sunlight.
- Substrate solution should remain colourless until added to the plate.
- To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.

12) PRECAUTIONS

All test components of human source were tested against HIV-Ab, HCV-Ab and HBsAg, and were found negative.

Nevertheless, they should be handled and disposed as if they were infectious. Liquid reagents contain $\leq 0.1\%$ Proclin 950 as preservative, which is not toxic in concentrations used in this kit. It may cause allergic skin reactions – avoid contact with skin or eyes.

- Do not pipette by mouth.
- Do not eat, drink, smoke or apply cosmetics where reagents are used.
- Avoid all contact with the reagents by using gloves.
- Sulfuric acid is irritating to eyes and skin. Flush with water if contact occurs.

13) LITERATURE

1. Utility of the Amino-Terminal Fragment of Pro Brain Natriuretic Peptide in Plasma. For the Evaluation of Cardiac Dysfunction in elderly Patients in Primary Health Care. Alehagen U et al., Clin Chem, 49:8; 1337-1346 (2003).
2. Head-to-head comparison of N-terminal pro-brain natriuretic peptide, brain natriuretic peptide and N-terminal pro-atrial natriuretic peptide in diagnosing left ventricular dysfunction. Hammerer-Lercher et al., Clin Chim Acta, 310(2),193-197 (2001).
3. Prognostic evaluation of neurohumoral plasma levels before and during beta-blocker therapy in advanced left ventricular dysfunction. Stanek B. et al., JACC, 38, 436-442 (2001).
4. Pulmonary arterial hypertension in rheumatic mitral stenosis: does it affect right ventricular function and outcome after mitral valve replacement? Pande S et al., Interact CardioVasc Thorac Surg, 9: 421-425 (2009).
5. Prognostic impact of body mass index in patients undergoing coronary artery bypass surgery. Sung SH et al., Heart, 97: 648-654 (2011).
6. Effect of piboserod, a 5-HT4 serotonin receptor antagonist, on left ventricular function in patients with symptomatic heart failure. Kjekshus JK et al., Eur J Heart Fail, 11: 771-778 (2009).
7. Plasma pro-B-type natriuretic peptide in the general population: screening for left ventricular hypertrophy and systolic dysfunction. Goetze JP et al., Eur Heart J, 27: 3004-3010 (2006).

1) EINLEITUNG

BNP wird hauptsächlich durch Myokard als Reaktion auf Volumenüberladung und erhöhtem Fülldruck synthetisiert. Reifes BNP besteht nach Verlust der Signalsequenz aus 108 Aminosäuren (proBNP oder BNP-108). Es wird in das physiologisch aktive BNP-32, weitere C-terminalen Fragmente (vgl.

http://www.uniprot.org/uniprot/P16860#PRO_000001532), sowie ein physiologisch inaktives N-terminales Peptid aus Aminosäuren 1-76 gespalten, das weiter proteolytisch abgebaut wird. Die BNP-Fragmente in der Zirkulation sind daher sehr heterogen.

BNP spielt eine Schlüsselrolle in der kardiovaskulären Homöostase mit biologischen Wirkungen einschließlich Natriurese, Diurese, Vasorelaxation, und der Hemmung von Renin und Aldosteron-Sekretion. Eine hohe Konzentration von BNP im Blut ist ein Hinweis auf Herzinsuffizienz.

Interessensgebiete:

- Herzinsuffizienz, akuter Myokardinfarkt (linksventrikuläre Dysfunktion)
- Niereninsuffizienz
- Adipositas und Diabetes
- Verschiedene Formen der sekundären Hypertonie
- Therapiebeobachtung von Patienten mit Herzinsuffizienz

2) INHALT DES KITS

CONT	KIT KOMPONENTEN	MENGE
PLATE	Polyklonaler Schaf anti NT-proBNP Antikörper, beschichtet auf Mikrotiterplattenstreifen im Streifenhalter. Verpackt im Aluminium Beutel mit Trockenmittel	12 x 8 Teste
WASHBUF	Waschpuffer, 20fach Konzentrat, durchsichtiger Verschluss	1 x 50 ml
STD	Standard, synthetisches humanes NT-proBNP (0, 10, 40, 160, 640 pmol/l), weißer Verschluss, lyophilisiert,	5 Fläschchen,
CTRL	Kontrollen, synthetisches humanes NT-proBNP, gelber Verschluss, lyophilisiert, genaue Konzentrationen nach Rekonstitution siehe Etiketten	2 Fläschchen,
CONJ	Konjugat (Schaf anti human NT-proBNP-HRPO), rot gefärbt, brauner Verschluss, gebrauchsfertig	1 x 22 ml
SUB	Substrat (TMB Lösung), blauer Verschluss, gebrauchsfertig	1 x 22 ml
STOP	Stopp Lösung, H ₂ SO ₄ , weißer Verschluss, gebrauchsfertig	1 x 7 ml

3) ZUSÄTZLICHES MATERIAL IM KIT

- 1 selbstklebende Abdeckfolie
- QC Protokoll
- Protokoll Blatt
- Arbeitsanleitung (Beipacktext)

4) ZUSÄTZLICH BENÖTIGTES MATERIAL UND GERÄTE

- Kalibrierte Präzisionspipetten für 50-500 µl inkl. Pipettenspitzen
- Mikrotiterplattenphotometer mit 450 nm Filter (630 nm Referenz optional)
- Millimeterpapier oder Software
- Mikrotiterplatten Wascher wird empfohlen, Alternativen: Mehrkanalpipetten oder Dispenser
- Kühlschrank mit 4°C (2-8°C)
- Destilliertes oder deionisiertes Wasser

5) REAGENZIEN UND PROBENVORBEREITUNG

Alle Reagenzien des Kits sind bei 4°C (2-8°C) bis zum Ablaufdatum (siehe Etikett des Reagenz) stabil.

Probenvorbereitung:

Abnahme von venösem Blut mittels standardisierter Blutabnahmeröhrchen zur Gewinnung von Serum oder Plasma. NT-proBNP ist im Vollblut für einige Stunden bei Raumtemperatur (18-26°C) stabil. Trotzdem wird die Zentrifugation des Blutes für 20 Minuten bei 2000 g, bevorzugt bei 4°C (2-8°C), so schnell wie möglich empfohlen. Serum oder EDTA-Plasma kann bis zu 2 Tage bei 4°C (2-8°C) aufbewahrt werden. Für Langzeit Lagerung sollen die Proben aliquotiert und bei -25°C

oder tiefer gelagert werden. Bis zu 3 Frier/Tau-Zyklen verändern die Messwerte nicht. Lipämische und hämolytische Proben können falsche Ergebnisse liefern. Proben vor Verwendung gut mischen. Wir empfehlen Doppelbestimmung für alle Werte. Proben mit Werten über STD5 (640 pmol/l) können mit STD1 oder NT-proBNP negativem human Serum verdünnt werden.

Nähere Informationen zur Probenstabilität finden Sie auf unserer Website www.bmgrp.com (s. Validation Data) oder kontaktieren Sie unseren Kundenservice per Email an info@bmgrp.com oder Tel. unter +43/ 1/ 29107-45.

Rekonstitution/Handhabung:

- **STD (Standard):** Pipettieren Sie 500 µl destilliertes oder deionisiertes Wasser in jeden Standard. Lassen Sie das Lyophilisat bei Raumtemperatur (18-26°C) 10 Minuten lösen. Gut mischen. Genaue Konzentration nach Rekonstitution siehe Etikett. Der rekonstituierte Standard ist bei -25°C oder niedriger bis zum Ablaufdatum haltbar. Vermeiden Sie Frier/Tau Zyklen.
- **CTRL (Kontrolle):** Pipettieren Sie 500 µl destilliertes oder deionisiertes Wasser in die Kontrollen. Lassen Sie das Lyophilisat bei Raumtemperatur (18-26°C) 10 Minuten lösen. Gut mischen. Genaue Konzentration nach Rekonstitution siehe Etikett. Die rekonstituierte Kontrolle ist bei -25°C oder niedriger bis zum Ablaufdatum haltbar. Vermeiden Sie Frier/Tau Zyklen.
- **WASHBUF (Waschpuffer):** Das mitgelieferte Konzentrat wird 1:20 (1+19) verdünnt, zB. 50 ml WASHBUF + 950 ml destilliertes Wasser. Kristalle im Pufferkonzentrat lösen sich bei Raumtemperatur auf. Der unverdünnte WASHBUF ist bei 4°C (2-8°C) bis zum Ablaufdatum haltbar. Der verdünnte WASHBUF ist bei 4°C (2-8°C) bis zu einem Monat haltbar. Im Testsystem darf nur verdünnter WASHBUF verwendet werden.

6) TESTPRINZIP

Siehe Kapitel 6) PRINCIPLE OF THE ASSAY im englischen Teil des Beipacktextes.

7) TESTPROTOKOLL

Im Test dürfen nur Reagenzien und Proben verwendet werden, welche Raumtemperatur (18-26°C) aufweisen.

Markieren Sie die Positionen für STD/PROBE/CTRL (Standard/Probe/Kontrolle) am Protokollblatt.

Nehmen Sie die benötigten Mikrotiterstreifen aus dem Aluminium Beutel. Nicht verwendete Mikrotiterstreifen können mit Trockenmittel im Aluminium Beutel bei 4°C (2-8°C) bis zum angegebenen Ablaufdatum gelagert werden.

1. Pipettieren Sie 50 µl STD/PROBE/CTRL (Standard/Probe/Kontrolle) in Doppelbestimmung in die entsprechenden Wells.
2. Pipettieren Sie 200 µl CONJ (Konjugat) in alle Wells. Gut mischen.
3. Streifen abdecken und 3 Stunden bei Raumtemperatur (18-26°C) inkubieren.
4. Inhalt der Wells verwerfen und 5x mit 300 µl verdünntem WASHBUF (Waschpuffer) waschen. Nach dem letzten Waschschritt Reste von Waschpuffer durch Ausklopfen auf saugfähigem Papier entfernen.
5. Pipettieren Sie 200 µl SUB (Substrat) in alle Wells. Gut mischen.
6. 30 min bei Raumtemperatur (18-26°C) im Dunkeln inkubieren.
7. Pipettieren Sie 50 µl STOP (Stopplösung) in alle Wells. Gut mischen.
8. Extinktion unmittelbar bei 450 nm messen, mit 630 nm als Referenz, falls möglich.

8) BERECHNUNG DER ERGEBNISSE

Messen Sie die optische Dichte (OD) von allen Wells mit einem 450 nm Filter (Referenz 630 nm). Konstruieren Sie eine Standardkurve aus den OD Werten der STD unter Verwendung von kommerziell erhältlichem Millimeterpapier oder einer geeigneten Software. Das Testsystem wurde mit einem 4 Parameter Algorithmus evaluiert. Andere Auswerte Algorithmen müssen vom Verwender evaluiert werden. Die Konzentration der Proben wird aus der Standardkurve abgelesen. Proben mit Ergebnissen über dem höchsten Standard (STD5, 640 pmol/l) können mit STD1 oder niedrig messendem Serum verdünnt und erneut getestet werden. Eventuelle weitere Verdünnungen müssen bei der Berechnung der Probenwerte berücksichtigt werden.

Typische STD-Kurve:

Siehe Kapitel 8) CALCULATION OF RESULTS im englischen Teil des Beipacktextes.

Auf dem beigeckten QC Protokoll sind die Resultate der QC Freigabe der jeweiligen Kit Lot vermerkt. Vom Verwender erhaltenen Daten der OD können abweichend sein, bedingt durch verschiedene Einflüsse und/oder dem normalen Signalverlust des Kits während der Laufzeit. Dieser mögliche Signalverlust hat keinen Einfluss auf die Gültigkeit der

Resultate, so lange die OD von STD5 den Wert 1,50 oder höher erreicht und der Wert der CTRL im gültigen Bereich ist (Bereich siehe Etikett).

9) TESTMERKMALE

Methode:	Sandwich ELISA, HRP/TMB, 12x8-well Streifen		
Antibodies/Standard:	Beschichtungsantikörper: polyklonal Schaf anti human NT-proBNP (aa 31-57) Detektionsantikörper: polyklonal Schaf anti human NT-proBNP (aa 8-29), HRPO konjugiert. Standard material: synthetisches human NT-proBNP (1-76).		
Probentyp:	Human serum, EDTA plasma		
Standardbereich:	0 bis 640 pmol/l (0, 10, 40, 160, 640)		
Umrechnungsfaktor pg/ml zu pmol/l:	1 pmol/l = 8,475 pg/ml Bezugnehmend auf NT-proBNP (1-76) Standardmaterial		
Probenvolumen:	50 µl / well		
Inkubationszeiten:	3 h / 30 min		
Sensitivität:	LOD (0pmol/l + 3SD): 3,0 pmol/l; LLOQ: 3,3 pmol/l		
Spezifität	Dieser Assay erkennt endogens und rekombinantes humanes NT-proBNP (1-76).		
Präzision:	Intra-assay (n=3) ≤ 4% , Inter-assay (n=8) ≤ 7%		
Wiederfindung (durchschnittliche R nach spike mit rek. NT-proBNP)	Spike 80 pmol/l	Spike 320 pmol/l	
	Serum (n=4)	99%	108%
	EDTA Plasma (n=4)	94%	93%
Verdünnungslinearität (durchschnittliche NT-proBNP Werte nach 1+1 VD)	Serum (rek. NT-proBNP): 117%	Serum (endog. NT-proBNP): 82%	
	EDTA plasma (rek. NT-proBNP): 84%	EDTA plasma (endog. NT-proBNP): 84%	
Werte von anscheinend gesunden Spendern:	Median (Serum, n=70): 5,8 pmol/l Median (EDTA plasma, n=28): 5,6 pmol/l Jedes Labor sollte den Normalbereich für seine Proben evaluieren. Wechseln Sie nicht die Probenmatrix innerhalb einer Studie.		

Nähere Informationen zu den Testmerkmalen finden Sie auf unserer Website unter www.bmgrp.com
(s. Validation Data) oder kontaktieren Sie unseren Kundenservice per Email an info@bmgrp.com oder telefonisch unter +43/ 1/ 29107-45.

10) PRÄZISION

Experiment:

Intra-assay: 2 Proben wurden 3 Mal in Doppelbestimmung in 1 Test von 1 Operator getestet.

Inter-assay: 2 Proben wurden 8 Mal in Doppelbestimmung in 2 Tests von unterschiedlichen Operatoren getestet.

Intra-assay (n=3)	Probe 1	Probe 2	Inter-assay (n=8)	Probe 1	Probe 2
Durchschnitt (pmol/l)	60,2	35,2	Durchschnitt (pmol/l)	52,1	108,1
SD (pmol/l)	2,0	0,9	SD (pmol/l)	1,7	7,9
VK (%)	4	3	VK (%)	3	7

11) TECHNISCHE MERKMALE

- Reagenzien von verschiedenen Lots oder Tests dürfen nicht gemischt werden.
- Stöpsel und Verschlüsse von verschiedenen Reagenzien dürfen nicht vertauscht werden.
- Abgelaufene Reagenzien dürfen nicht verwendet werden. Reagenzien sind vor direktem Sonnenlicht zu schützen.
- Substratlösung muss vor Verwendung farblos sein.
- Mikrotiterstreifen müssen bei den Inkubationen mit Abdeckfolie abgedeckt sein.

12) VORSICHTSMASSNAHMEN

Alle Bestandteile humanen Ursprungs wurden auf HIV-Ak, HCV-Ak und HBsAg getestet und negativ gefunden. Trotzdem sollten die Reagenzien als potentiell infektiös behandelt werden. Die flüssigen Reagenzien enthalten ≤0,1% Proclin 950 als Konservierungsmittel. Vermeiden Sie Kontakt mit Augen, Haut und Schleimhaut. Proclin 300 ist in der verwendeten Konzentration nicht toxisch. Allergische Reaktionen sind möglich.

- Nicht mit dem Mund pipettieren.
- Nicht Rauchen, Essen, Trinken oder Kosmetika benutzen während der Verwendung der Testreagenzien.
- Verwenden Sie Handschuhe, Schutzbrille und Laborkleidung während der Testdurchführung.
- Schwefelsäure reizt Augen und Haut. Bei Berührung gründlich mit Wasser spülen.

13) LITERATUR

Siehe Kapitel 13) *LITERATURE* im englischen Teil des Beipacktextes.

SYMBOLS



Expiry date / Verfallsdatum / Date de péremption / Data di scadenza / Fecha de caducidad / Data de validade / Uiterste gebruiksdatum / Udløbsdato / Utgångsdatum / Termin Ważności / Lejárat idő / Doba expirácie / Doba expirace



Consider instructions for use / Bitte Gebrauchsanweisung beachten / Consultez la notice d'utilisation / Consultare le istruzioni per l'uso / Consulte las instrucciones de utilización / Consulte as instruções de utilização / Raadpleeg de gebruiksaanwijzing / Se brugsanvisningen / Lás anvisningarna före användning / Proszę przeczytać instrukcję wykonania / Vegyük figyelembe a használati utasításban foglaltakat / Postupujte podľa pokynov na použitie / Postupujte dle návodu k použití



In vitro Diagnostic Medical Device (for in Vitro Diagnostic Use) / In vitro Diagnostikum (zur In-vitro-Diagnostik) / Dispositif médical de diagnostic in vitro (Pour usage diagnostique in vitro) / Dispositivo medico per diagnostica in vitro (per uso diagnostico in vitro) / Dispositivo médico de diagnóstico in vitro (para uso diagnóstico in vitro) / Dispositivo médico para diagnóstico in vitro (Para utilização de diagnóstico "in vitro") / Medicisch hulpmiddel voor diagnostiek in vitro (Voor diagnostisch gebruik in vitro) / Medicinsk udstyr til in vitro-diagnostik (Udelukkende til in vitro diagnostisk anvendelse) / Medicinteknisk produkt avsedd för in vitro-diagnostik (För in vitro-diagnostiskt bruk) / Wyrob medyczny do Diagnostyki In Vitro / In vitro orvostudományi termék / In vitro diagnostický zdravotnícky materiál (určené pre diagnostiku „in vitro“) / In vitro diagnostický zdravotnícky materiál (určeno pro diagnostiku „in vitro“)



Lot-Batch Number / Charge-Chargennummer / Lot-Code du lot / Lotto-Numero di lotto / Lote-Código de lote / Lote-Código do lote / Lot-Partijnummer / Lot-Batchkode / Lot-Satskod / Numer serii / Lot-Batch szám / Číslo šárže / Číslo šárže



Manufactured by / Hergestellt von / Fabriqué par / Prodotto da / Fabricado por / Fabricado por / Vervaardigd door / Fabrikation af / Tillverkad av / Wyprodukowane pr / Gyártotta / Vyrobené / Vyrobeno



Catalogue Number / Bestellnummer / Numéro de référence / Numero di riferimento / Número de referencia / Número de referência / Referentienummer / Referencenummer / Katalognummer / Numer katalogowy / Katalógusszám / Katalógové číslo / Katalogové číslo



Store at between / Lagerung bei zwischen / Conserver à entre / Conservare a tra / Conservar a temp. entre / Armazene a entre / Bewaar bij tussen / Opbevares mellem / Förvaras vid / Przechowywać w / Tároljuk között / Skladujte v rozsahu / Skladujte v rozmezí



Contains sufficient for x tests / Inhalt ausreichend für x Tests / Contient suffisant pour x tests / Contenuto sufficiente per x test / Contiene suficiente para x pruebas / Contém suficiente para x testes / Bevat voldoende voor x bepalingen / Indeholder tilstrækkeligt til x prøver / Innehåller räcker till x analyser / Zawartość na x testów / Tartalma X teszt elvégzésére elegendő / Obsahuje materiál pre x testov / Obsahuje materiál pro x testů

SK-1204 NT-proBNP

ASSAY PROTOCOL AND CHECKLIST

PREPARATION OF REAGENTS:

- Bring all reagents to room temperature (18-26°C).
- Prepare reagents and samples as instructed.
- Bring unused and prepared components to the storage temperature mentioned in the package insert.
- Take microtiter strips out of the alu bag and mark positions on the protocol sheet.

TEST PROCEDURE:

- Step1) Add 50 µl STD/ SAMPLE/CTRL (standard/sample/control) in duplicate into respective well.
- Step 2) Add 200 µl CONJ (Conjugate) into each well. Swirl gently.
- Step 3) Cover tightly and incubate for 3 hours at room temperature (18-26°C).**
- Step 4) Aspirate and wash wells 5x with 300 µl diluted WASHBUF (Wash buffer). Remove remaining buffer by strongly tapping plate against paper towel after last wash.
- Step 5) Add 200 µl SUB (Substrate) into each well. Swirl gently.
- Step 6) Incubate for 30 minutes at room temperature (18-26°C) in the dark.**
- Step 7) Add 50 µl STOP (Stop solution) into each well. Swirl gently.
- Step 8) Read Optical Density at 450 nm with reference 630 nm, if available.